# Protocol for Real-Time RT-PCR assay with Lyoph P&P Monkeypox F3L (adapted to TaqMan platform)

## For RUO (Research Use Only)

#### Vial content

Lyoph P&P Monkeypox F3L contains lyophilized primers and probe (P&P) for detection of monkeypox virus (MPXV). It must be used in combination with a generic RT-qPCR kit.

#### Design

RT-qPCR systems included in Lyoph P&P Monkeypox F3L:

 Monkeypox virus detection in rodents using real-time 3'-minor groove binder TaqMan assays on the Roche LightCycler. Kulesh et al., Laboratory Investigation, 2004. <u>DOI</u>: 10.1038/labinvest.3700143

#### Caution

Vials containing primers and probe mix (Lyoph P&P) must be stored at -20°C in the dark after reception. Stable 4 years under the described conditions.

#### Instructions

This protocol is adapted to the following kit and RT-qPCR platform:

- GoTaq® qPCR Master Mix (Promega)
- CFX (BIORAD) RT-qPCR thermal cycler

Use of other kits or RT-qPCR platforms may require adaptation of the protocol.

#### I. Rehydration of Lyoph-P&P

- Write the date on the vial before opening
- Resuspend Lyoph P&P as follows (these steps are critical to ensure adequate homogenization):
  - ✓ Add nuclease-free water or RE buffer as described in Table 1
  - ✓ Homogenize by pipetting 10 to 20 times the volume up and down in the glass vial
  - ✓ Incubate rehydrated P&P at room temperature for 10 min
  - ✓ Perform a second series of 10-times pipetting.

**Table 1.** Lyoph-P&P regeneration; the number of tests/vial is indicated on the Lyoph P&P vial label.

Packaging (number of tests/vial)	8+1	16+2	24+2	32+3	48+3	96+5
H₂O or RE buffer (μL)	63	126	182	245	357	707

# II. <u>Preparation of the PCR reaction mix (for GoTaq® qPCR Master Mix - Promega)</u>

**Table 2**. Preparation of the PCR mix. \*Rehydration performed as indicated in Table 1.

	Volume/PCR test (μL)
2X Master mix	13
Rehydrated Lyoph P&P*	7
Total (1 PCR test)	20
Template DNA (H₂O for Ctrl-)	5
Final volume	25

## III. Cycling program and RT-qPCR (for GoTaq® qPCR Master Mix - Promega)

95°C	2min	
95°C	15sec	45 cycles
60°C	45sec (plate read)	

Read: FAM-QSY

Rehydrated Lyoph-P&P can be aliquoted and stored at -20°C. Each aliquot can be thawed only once and then discarded.